

QUOTATION NOTICE

Sealed quotations are invited from the eligible firms for the supply and installation of Vacuum Pump in the Dept. of Environmental Sciences, University of Kerala, Kariavattom Campus.

Technical Specifications for Vacuum pump

Specification	Requirements
Power	220V/50hz
Maximum power (W)	80
Maximum vacuum	-650mm/Hg
Maximum flow rate	34 L/min
Motor Rotation	1450 RPM
Horse power	1/6 HP
Port. Thread	5/16 inch (8mm)
Noise level	50db
Net. Weight	5.2kg
Other requirements	Moisture Trap, VacuumRegulator, Overflow protection
Warranty	One year

Conditions

1. The rate quoted should be inclusive of Tax, transportation charges and any other charges.
Note: The University Issue Essentiality Certificate, so tax should be quoted Accordingly.
2. Complaints with the items supplied should be rectified /replaced to the satisfaction of the University by the suppliers at their own cost.
3. Payment will be made only after the supply of the items at our own site.
4. The undersigned reserves the right to reject /accept any quotation without assigning any reason.
5. The equipment should be installed in the Department of Environmental Sciences, University of Kerala.

The sealed quotations should reach **‘The Associate Professor and Head, Department of Environmental Sciences, University of Kerala, Kariavattom Campus P.O. Thiruvananthapuram-695 581, Kerala** on or before **28/11/2018 at 3.00 p.m.**