



DEPARTMENT OF BIOCHEMISTRY

(DST-FIST Sponsored Department)
UNIVERSITY OF KERALA

(Established as University of Travancore by the Travancore University act in 1937 and reconstituted as University of Kerala by the Kerala University act of 1957 and presently governed by the Kerala University act of 1974 passed by the Kerala state Legislative Assembly)

(Re – accredited by NAAC with 'A⁺⁺' Grade)



KARIAVATTOM
THIRUVANANTHAPURAM – 695581
PHONE: 0471 2308078

Email: biochemistry@keralauniversity.ac.in

22.11.2023

QUOTATION NOTICE

(BCH/SPL FND-LAB DVLPMNT/NP-01/2023-24)

Sealed two bid quotations are invited for the supply of **Fume Hood With Two Blower** with relevant specifications with on or before **04/12/2023** at **4pm**. Sealed quotations should be addressed to Professor & Head, Department of Biochemistry, University of Kerala, Thiruvananthapuram- 695581 with superscription. Quotation for the supply of **Fume Hood With Two Blower**. The quotation will be opened on **05/12/2023** at 2pm in the presence of such tenderers or their authorized representatives who may be present at that time. The rate quoted shall be valid for a period of one year. Quotations should also include all the supportive work for installation, loading and unloading of the equipment. Bidders have to note Appendix I - Technical Specifications and Appendix II Terms and Conditions and should attach the relevant documents along with the quotation.

Name of the Equipment: Fume Hood with TWO Blower

Appendix I – Technical Specifications


Fume Hood with TWO Blower

- Overall Dimension : 5' x 3' x 5' (L x W x H)
- Inner Working Area : 5'W x 2.4'L x 3.0'H

- Cabinet made of GI Epoxy powder coated with Inner working area FRP Coated
- Blower Air flow: Designed for 0.45 to 0.65 m/s
- Inner Areas are FRP coated
- Sash opening space-3.5 feet
- LED lamp
- Front Door:6 MM thickness
- Polycarbonate/Glass, vertically sliding,
- Balanced with counter weight.
- LPG Gas Control valve – 1No
- N2 Gas valve – 1 No
- Z Black Granite(19mm) Top work table
- Extra power outlet switch
- Power supply : 230 V AC
- Noise level : 65 +/- 5 db

Appendix II – Terms and Conditions

- Delivery within 4-6 weeks from the date of receipt of our order
- 2-3 years Warranty


Professor & Head
Department of Biochemistry
University of Kerala, Kariavattom
Thiruvananthapuram - 695 581

