

**Department of Physics
University of Kerala,
Thiruvananthapuram-695581**

Quotation Notice

17/01/2021

Department of Physics, University of Kerala, Thiruvananthapuram invites quotations for the purchase of Probe sonicator. The specifications required are given below.

Required specifications for the Probe sonicator

- Frequency: 20-25 kHz automatic tracking
- Process Capacity: 0.5-500ml or better
- Probe Dia (Titanium): 12 - 20 mm
- Display Function: Temperature, Power, Time
- Temperature range (°C): 0-99°C or higher
- Timer (Selectable): should be provided
- Overload Protection: Should be provided
- Output Power: 6.5-650W(should be variable)
- Power Supply : AC220V/50Hz
- Approximate Dimensions (LxWxH) mm: 500x400x200

Please quote the lowest price for the above item.

The quotations should reach **"The Head, Department of Physics, University of Kerala, Kariavattom, Thiruvananthapuram, Pin- 695 581, Kerala"** or o before **28/01/2022 3.00 PM.**

Sibi K. S.

