

**Centre for Nanoscience and Nanotechnology, University of Kerala,
Kariavattom, Thiruvananthapuram-695581
Ph.No. 0471 2308863**

Quotation Notice

No. KU/CNST/QN-3/2013-14

Quotations are invited for the supply of one number of table top **high temperature magnetic stirrer** with PT 100 sensor probe of make Glassco-Lobalife/Amkette Industries/IKA/Benchmark Scientific with the following specifications.

Name of item	Description	
Table top high temperature magnetic stirrer with PT 100 sensor probe	Table top high temperature magnetic stirrer with PT 100 sensor probe	
	Stirring capacity	1 Litre, maximum
	Heating Power	500 W (approx.)
	Temperature range	Room temperature to 350°C (approx.)
	Speed	50 to 1700 rpm (approx.)
	Heat control accuracy	± 2°C or better

Kindly quote your lowest price for the above item. Sealed quotations should reach the “Honorary Director, Centre for Nanoscience and Nanotechnology, University of Kerala, Kariavattom, Thiruvananthapuram-695581 on or before 05-09-2013, 3:00 pm.

S/d

02-08-2013

Honorary Director