

Department of Chemistry, University of Kerala,
Kariavattom, Thiruvananthapuram – 695581
Phone: 0471- 2308682

Quotation Notice

No.Dc/ST/Eq/60/13-14

Quotations are invited for the following items in the Department of Chemistry, University of Kerala, Kariavattom, Thiruvananthapuram

Item1: Autoclave (Vertical): Double walled made of stainless steel; with dimensions 350 x 550 mm; Watts 3.0 kW; Approximate Volume 53 L; preferably with Digital Temperature Controller cum Indicator , Digital Timer and Auto Pressure cut - off system

Item2: Laminar Air Flow Cabinet (Horizontal): With outer dimensions (W x D x H) 40” x 40” 60 ½; Size of HEPA Filter (“CLASS 100”) 3’ x 2’ with an efficiency of 99% or better; Work table made of stainless steel; Air blow velocity of 90 ft/min or better.

Item3: Bacteriological Incubator: Bottom Heated with Chamber Size (L x B x H) 450x450x450 mm; Number of Trays 2; Watts 250 kW; Approximate Volume 95 L; with Digital Temperature Controller, and working temperature ambient + 5°C to 80°C with an accuracy of +/-1°C and without thermometer and Air Circulation Fan.

The firms who wish to supply the items are requested to submit the quotation in sealed cover to the undersigned on or before 24/03/2014.

Kariavattom
18/03/2014

Associate Professor & Head